



HYGITEST Ltd. –Sofia, Bulgaria
15, Acad. Iv. Evst. Geshov, Blvd.
Fax/Phone: (0359 2) 954-11 19
www.hygitest.com :e-mail : info@hygitest.com



DQ 07_00-12

CERTIFICATE OF QUALITY
No. SS6 – 4 /2017
SILICA GEL SAMPLING TUBES

Characteristics	Value
Type of the sorbent tubes	Silica Gel + H ₂ SO ₄
Catalogue No.	ST 18-31
Lot #	0720
Expiration date	08.2019
Design parameters	
Overall length of the glass (sealed) tube, mm	70 ± 2
Outside diameter, mm	6.0 ± 0.2
Internal diameter, mm	4.0 ± 0.2
Diameter of the tips, mm	2.0 ± 0.3
Weight of the backup sorbent layer, mg	100 ± 5
Weight of the front sorbent layer, mg	200 ± 10
Characteristics of materials	
Glass tubes	Neutral glass
Type of glass wool separators	Silanized
Filling material	Silica Gel treated with sulfuric acid to meet the specifications of the NIOSH methods 6015 and 6016.

Sofia, 2017

Produce Manager*
 /Eng.  D. Vasileva/